

Accreditation Scope

 Test TS EN ISO/IEC 17025 AB-0177-T	BAYLAN Ölçü Aletleri Sanayi Ve Ticaret Ltd. Şti.	
	Accreditation Number: AB-0177-T Revision Number: 00 Date: 05 March 2008	
Test Laboratory		
Address: Atatürk Organize Sanayi Bölgesi 10046 Sok. No:14 Çiğli İZMİR / TÜRKİYE		Phone : 0 232 376 77 57 Fax : 0 232 376 77 59 E-Mail : baylan.lab@gmail.com Website: www.baylansayaclari.com

Tested Materials / Products	Name of Test	Testing Method (National, International standards, in house methods)
Cold Water Meters	Pressure Loss Test	TS 10021 Article 4 ISO 4064-3:2005 Article 7 OIML 49-2:E06 Article 6.7 Directive for Cold Water Meters 75/33 EEC Article 3.4 EN 14154-3:2005 Article 5.11
	Leakage Test	TS 10021 Article 3 ISO 4064-3:2005 Article 6 OIML 49-2:E06 Article 6.2 Directive for Cold Water Meters 75/33 ECC Article 3.3 EN 14154-3:2005 Article 5.2
	Test for determination of error curves according to flow rates	TS 10021 Article 2 ISO 4064-3:2005 Article 5 OIML 49-2:E06 Article 6.3 Directive for Cold Water Meters 75/33 ECC Article 2 EN 14154-3:2005 Article 5.3
	Accelerated Life Test (Life tests with continuous flows and with interrupted flows)	TS 10021 Article 5 ISO 4064-3:2005 Article 8.1 OIML 49-2:E06 Article 6.9 Directive for Cold Water Meters 75/33 ECC Article 5 EN 14154-3:2005 Article 5.13
	Reverse Flow Test	ISO 4064-3:2005 Article 5.12 OIML 49-2:E06 Article 5.11 EN 14154-3:2005 Article 5.12
	Zero Flow Test	EN 14154-3:2005 Article 5.5

End of Scope



Emre SEZER
Acting Secretary General